TEL: 55-51-3323 3600 FAX: 55-51-3323 3644 info@novus.com.br

DECLARATION OF CONFORMITY

1. Declaration of Conformity

This Declaration of Conformity is made against the Electromagnetic Compatibility Directive (89/336/EEC).

2. Product model: IrLink3

3. Product description

The *IrLink3* is a USB-to-Infrared communication interface used for configuring parameters and retrieving data in data logger instruments carrying infrared communication port. Its power is provided by the USB port of the computer.

4. Conformity

The *IrLink3* conforms to the following standards:

- EMC: EN 61326-1:1997 and EN 61326-1/A1:1998

Emission:	Standard	Description	
	CISPR22	Irradiated emission	
Immunity:	Standard	Description	Severity level
illilliulity.			
	EN61000-4-2	Electrostatic Discharge test	CD 4 kV
			AD 8 kV
	EN61000-4-3	Radiated Susceptibility test	3 V/m
	EN61000-4-4	Electrical Fast Transients	Not applicable
	EN61000-4-5	Voltage surge	Not applicable
	EN61000-4-6	Conducted disturbances,	Not applicable
		Induced by radio frequency fields	
	EN61000-4-8	Magnetic Field Immunity	30 A/m, 50 Hz

- Safety: EN61010-1:1993 and EN61010-1/A2:1995

5. Manufacturer:

NOVUS PRODUTOS ELETRONICOS LTDA Rua Álvaro Chaves, 167 Porto Alegre, RS, BRAZIL 90220-040

Porto Alegre, 05/MAR/2010

Miguel Freehung

Miguel Fachin Jr. R&D Director